

VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 5464(1-3) Exp date: 2020-Mar-31

Acculite® Clia	54641		54642		54643	
	Mean	Range	Mean	Range	Mean	Range
Cancer Markers						
Ferritin in ng/ml	31.00	20.77 - 41.23	245.00	164.15 - 325.85	660.00	442.20 - >800
CEA in ng/ml	2.80	1.88 - 3.72	24.00	16.08 - 31.92	80.00	53.60 - 106.40
CEA-NG in ng/ml	1.80	1.21 - 2.39	23.00	15.41 - 30.59	80.00	53.60 - 106.40
AFP in ng/ml	6.70	4.49 - 8.91	70.00	46.90 - 93.10	215.00	144.05 - 285.95
PSA in ng/ml	<0.5	N/A	6.10	4.09 - 8.11	28.50	19.10 - 37.91
fPSA in ng/ml	<0.5	N/A	2.80	1.88 - 3.72	>10	N/A
CA-125 in U/ml	21.11	14.14 - 28.08	56.56	37.90 - 75.22	165.81	111.09 - 220.53
CA-19-9 in U/ml	11.00	7.37 - 14.63	25.00	16.75 - 33.25	76.00	50.92 - 101.08
CA 15-3 in U/ml	14.50	9.72 - 19.29	39.00	26.13 - 51.87	75.00	50.25 - 99.75
PRL in ng/ml	8.60	5.76 - 11.44	33.75	22.61 - 44.89	>100	N/A
HCG in mIU/ml	5.10	3.42 - 6.78	52.10	34.91 - 69.29	195.40	130.92 - >250
HCG-XR in mIU/ml	2.05	1.37 - 2.73	51.50	34.51 - 68.50	215.00	144.05 - 285.95
TG in ng/ml	6.10	4.09 - 8.11	23.00	15.41 - 30.59	51.00	34.17 - 67.83

DOCUMENT HISTORY

 PREPARED BY: *MSW* DEPT: QC
 APPROVED BY: *Fred Lopez* DEPT: Administration
 REVISION: 0

 VERIFIED BY: *Ashatola* DEPT: QA
 EFFECTIVE DATE: 2017-Feb-28
 DCO: N/A