

VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 5464(1-3)
Exp date: 2020-Mar-31

AccuBind® ELISA	54641		54642		54643	
	Mean	Range	Mean	Range	Mean	Range
Tumor Markers						
Ferritin in ng/ml	29.70	19.90 - 39.50	232.30	155.64 - 308.96	606.51	406.36 - >800
CEA in ng/ml	3.20	2.14 - 4.26	25.50	17.09 - 33.92	88.00	58.96 - 117.04
CEA-NG in ng/ml	2.50	1.68 - 3.33	24.00	16.08 - 31.92	88.00	58.96 - 117.04
AFP in ng/ml	6.40	4.29 - 8.51	72.00	48.24 - 95.76	218.00	146.06 - 289.94
PSA in ng/ml	<0.5	N/A	6.20	4.15 - 8.25	31.00	20.77 - 41.23
fPSA in ng/ml	<0.5	N/A	2.40	1.61 - 3.19	>10	N/A
CA-125 in U/ml	23.80	15.95 - 31.65	61.32	41.08 - 81.56	175.80	117.79 - 233.81
CA-19-9 in U/ml	16.00	10.72 - 21.28	30.00	20.10 - 39.90	80.00	53.60 - 106.40
CA 15-3 in U/ml	22.00	14.74 - 29.26	40.00	26.80 - 53.20	90.00	60.30 - 119.70
PRL in ng/ml	9.30	6.23 - 12.37	32.50	21.78 - 43.23	>100	N/A
HCG in mIU/ml	6.00	4.02 - 7.98	50.00	33.50 - 66.50	200.00	134 - >250
HCG-XR in mIU/ml	3.95	2.65 - 5.25	55.47	37.16 - 73.78	226.92	152.04 - 301.80
TG in ng/ml	5.80	3.89 - 7.71	21.00	14.07 - 27.93	48.00	32.16 - 63.84

DOCUMENT HISTORY

 PREPARED BY: NSW DEPT: QC

 VERIFIED BY: Ashatola DEPT: QA

 APPROVED BY: Fred Lopez DEPT: Administration

EFFECTIVE DATE: 2017-Feb-28

REVISION: 0

DCO: N/A