

VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT 2367(1-3)				Exp: 2020-05-31		
AccuLite® CLIA	23671		23672		23673	
	Mean	Range	Mean	Range	Mean	Range
Cardiac Markers						
CKMB in ng/ml	2.19	1.47-2.91	9.25	6.20-12.31	47.57	31.87-63.27
hsCRP in µg/ml	0.87	0.58-1.17	2.07	1.39-2.76	5.88	3.94-7.82
Myoglobin in ng/ml	93.37	62.56-124.18	152.38	102.09-202.66	321.27	215.25->400
cTnl in ng/ml	<0.2		1.12	0.75-1.48	5.66	3.79-7.53

Note: Monobind evaluates levels 1, 2 and 3 in the Bio-Rad control.

It does not evaluate levels 1A, 1B and 1C due to minimal difference in value recognition.

DOCUMENT HISTORY

 PREPARED BY: *MSW* DEPT: QC

 VERIFIED BY: *A Shatola* DEPT: QA

 APPROVED BY: *Fred Lopez* DEPT: Administration

EFFECTIVE DATE: 2017-July-20

REVISION: 0

DCO: N/A