



VALUE ASSIGNMENT FOR BIO-RAD

MASTER LOT 5465(1-3) Exp date: 2020-07-31

Acculite® Clia	54651		54652		54653	
	Mean	Range	Mean	Range	Mean	Range
Tumor Marker						
Ferritin in ng/ml	32.84	22-43.68	282.91	189.55-376.27	720.03	482.42->800
CEA in ng/ml	2.78	1.86-3.70	21.10	14.14-28.06	86.78	58.14-115.41
CEA-NG in ng/ml	2.41	1.61-3.20	22.32	14.95-29.69	81.50	54.6-108.39
AFP in ng/ml	7.56	5.07-10.06	65.62	43.97-87.28	217.09	145.45-288.73
PSA in ng/ml	<0.5		6.16	4.12-8.19	27.90	18.69-37.11
fPSA in ng/ml	<0.5		2.41	1.61-3.20	>10	
CA-125 in U/ml	16.70	11.19-22.22	50.51	33.84-67.18	160.14	107.29-212.98
CA-19-9 in U/ml	11.31	7.58-15.04	25.74	17.25-34.23	59.85	40.10-79.6
CA 15-3 in U/ml	15.30	10.25-20.35	30.43	20.39-40.47	51.12	34.25-67.99
PRL in ng/ml	10.23	6.85-13.60	37.34	25.02-49.67	>100	
HCG in mIU/ml	5.22	3.50-6.94	49.58	33.22-65.94	208.39	139.62->250
HCG-XR in mIU/ml	3.96	2.65-5.27	47.78	32.01-63.55	199.95	133.97-265.93
TG in ng/ml	3.51	2.35-4.67	24.61	16.49-32.72	57.62	38.61-76.63

DOCUMENT HISTORY

PREPARED BY: MSW DEPT: QC  
 APPROVED BY: Fred Aguirre DEPT: Administration  
 REVISION: 0

VERIFIED BY: Ashatola DEPT: QA  
 EFFECTIVE DATE: 2017-Jun-01  
 DCO: N/A