

VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 4223(1-3) EXP: 2020-Jun-30

AccuLite® CLIA	42231		42232		42233	
	Mean	Range	Mean	Range	Mean	Range
Fertility						
Ferritin in ng/ml	7.32	4.90-9.73	30.95	20.74-41.16	70.54	47.26-93.81
AFP in ng/ml	19.63	13.15-26.11	73.9	49.51-98.29	309.43	207.32-411.54
LH in mIU/ml	133.12	89.19-177.04	49.09	32.89-65.29	33.32	22.32-44.31
hCG in mIU/ml	<2		20.28	13.59-26.97	18702	12530.34-24873.66
Testosterone in ng/ml	2.97	1.99-3.95	8.67	5.81-11.54	>10	
Progesterone in ng/ml	21.02	14.08-27.96	40.45	27.10-53.80	>60	
E2 in pg/ml	170.61	114.31-226.71	259.27	173.71-344.83	770.57	516.28-1024.86

DOCUMENT HISTORY

 PREPARED BY: *MSW* DEPT: QC

 VERIFIED BY: *A Shatola* DEPT: QA

 APPROVED BY: *Fahlgren* DEPT: Administration

EFFECTIVE DATE: 2017-Jun-14

REVISION: 0

DCO: N/A