



VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT 2368(1-3)				Exp: 2020-Aug-31		
AccuLite® CLIA	23681		23682		23683	
	Mean	Range	Mean	Range	Mean	Range
Cardiac Markers						
CKMB in ng/ml	2.28	1.53-3.03	13.01	8.72-17.30	67.48	45.21-89.75
hsCRP in µg/ml	0.81	0.54-1.08	2.1	1.40-2.79	6.72	4.50-8.94
Myoglobin in ng/ml	92.82	62.19-123.45	154.06	103.22-204.90	282.41	189.22-375.61
cTnl in ng/ml	<0.3		0.96	0.64-1.27	4.66	3.13-6.20

DOCUMENT HISTORY			
PREPARED BY: <u>MSW</u>	DEPT: QC	VERIFIED BY: <u>A Shatola</u>	DEPT: QA
APPROVED BY: <u>[Signature]</u>	DEPT: Administration	EFFECTIVE DATE: 2018-Feb-14	
REVISION: 1		DCO: 1299	