

VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 2526(1-3) EXP: 2020-Apr-30

AccuLite® CLIA	25261		25262		25263	
	Mean	Range	Mean	Range	Mean	Range
Lyphocheck Specialty IA						
SHBG in nmol/L	8.07	5.40-10.73	30.14	20.19-40.09	123.54	82.77-164.30
Procalcitonin (PCT) in ng/ml	0.67	0.45-0.89	1.66	1.11-2.21	16.97	11.37-22.57
Vitamin D Direct in ng/ml	12.65	8.47-16.82	22.71	15.21-30.20	63.94	42.84-85.04

DOCUMENT HISTORY

PREPARED BY: *MSW* DEPT: QC
 APPROVED BY: *Fred Lopez* DEPT: Administration
 REVISION: 0

VERIFIED BY: *Ashatola* DEPT: QA
 EFFECTIVE DATE: 2018-Mar-02
 DCO: N/A