

VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT 2396(1-3)			Exp: 2020-Apr-30			
AccuBind® ELISA	23961		23962		23963	
	Mean	Range	Mean	Range	Mean	Range
Tumor Markers						
Ferritin in ng/ml	28.35	18.95 - 37.75	375.26	251.42 - 499.10	>700	N/A
CEA in ng/ml	3.68	2.46 - 4.90	32.99	22.10 - 43.88	123.83	82.97 - 164.69
AFP in ng/ml	9.16	5.78 - 12.54	79.79	53.46 - 106.12	270.59	181.29 - 359.89
PSA in ng/ml	<0.5	N/A	9.57	6.41 - 12.73	36.89	24.72 - 49.06
fPSA in ng/ml	<0.3	N/A	3	2.01 - 3.99	>10	N/A
CA-125 in U/ml	29.93	20.05 - 39.81	66.77	44.73 - 88.81	181.39	121.53 - 241.25
CA-19-9 in U/ml	7.35	4.93 - 9.77	34.83	23.34 - 46.32	107.68	72.15 - 143.21
CEA Next Generation in ng/ml	2.51	1.68 - 3.34	28.58	19.15 - 38.01	96.04	64.35 - 127.73
PRL in ng/ml	11.05	7.4 - 14.7	36.01	24.13 - 47.89	>100	N/A
HCG in mIU/ml	2.69	1.77 - 3.61	32.26	21.61 - 42.91	159.06	106.57 - 211.55
TG in ng/ml	6.09	4.08 - 8.1	42.77	28.66 - 56.88	120.64	80.83 - 160.45
CA 15-3 in U/ml	16.48	11.04 - 21.92	31.34	21 - 41.68	51.78	34.69 - 68.87

DOCUMENT HISTORY			
PREPARED BY: <u>NSW</u>	DEPT: QC	VERIFIED BY: <u>AShatla</u>	DEPT: QA
APPROVED BY: <u>Fred Lopez</u>	DEPT: Administration	EFFECTIVE DATE: 2018-Apr-27	
REVISION: 0		DCO: N/A	

