



VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 4224(1-3) EXP: 2020-Dec-31

AccuLite® CLIA	42241		42242		42243	
	Mean	Range	Mean	Range	Mean	Range
Fertility						
Ferritin in ng/ml	8.77	5.88 - 11.66	46.99	31.48 - 62.50	96.46	64.63 - 128.29
AFP in ng/ml	22.84	13.98 - 31.70	80.3	53.80 - 106.80	303.52	203.36 - 403.68
LH in mIU/ml	133.8	89.65 - 177.96	46.72	31.30 - 62.14	32.5	21.77 - 43.22
hCG in mIU/ml	1.77	1.18 - 2.35	25.21	16.89 - 33.53	21994.67	14736.43 - 29252.91
Testosterone in ng/ml	0.28	0.19 - 0.37	3.43	2.30 - 4.56	7.27	4.87 - 9.67
Progesterone in ng/ml	20.96	14.04 - 27.87	44.28	29.67 - 58.90	>60	N/A
E2 in pg/ml	171.71	115.05 - 228.38	218.42	146.34 - 290.50	883.85	592.18 - 1175.52

DOCUMENT HISTORY

PREPARED BY: MSW DEPT: QC VERIFIED BY: AShatok DEPT: QA
APPROVED BY: Fred Lopez DEPT: Administration EFFECTIVE DATE: 2018-Mar-01
REVISION: 0 DCO: N/A