

VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT 6760(1-3)				Exp: 2021-Mar-31		
AccuLite® CLIA	67601		67602		67603	
	Mean	Range	Mean	Range	Mean	Range
Cardiac Markers						
CKMB in ng/ml	<5	n/a	12.62	8.46-16.78	67.94	45.52-90.36
hsCRP in µg/ml	0.73	0.49-0.97	2.01	1.35-2.68	6.01	4.03-8.0
Myoglobin in ng/ml	87.76	58.8-116.72	160.68	107.65-213.7	279.85	187.5-372.2

Note: Monobind evaluates levels 1, 2 and 3 in the Bio-Rad control.

It does not evaluate levels 1A, 1B and 1C due to minimal difference in value recognition.

DOCUMENT HISTORY

 PREPARED BY: *MSW* DEPT: QC

 VERIFIED BY: *AShatola* DEPT: QA

 APPROVED BY: *Fred Lopez* DEPT: Administration

EFFECTIVE DATE: 2018-May-15

REVISION: 0

DCO: N/A