



Testosterone Test System Product Code: 3725-300

1.0 INTRODUCTION

Intended Use: The Quantitative Determination of Total Testosterone Concentration in Human Serum or Plasma by a Microplate Enzyme Immunoassay, Colorimetric

2.0 SUMMARY AND EXPLANATION OF THE TEST

Testosterone, (17 β -Hydroxy-4-androstene-3-one), a C₁₉ steroid, is the most potent naturally secreted androgen.¹ In normal post pubertal males, testosterone is secreted primarily by the testes with only a small amount derived from peripheral conversion of 4-Androstene-3, 17-dione (ASD).² In adult women, it has been estimated that over 50% of serum testosterone is derived from peripheral conversion of ASD secreted by the adrenal and ovary, with the remainder from direct secretion of testosterone by these glands.

In the male, testosterone is mainly synthesized in the interstitial Leydig cells and the testis, and is regulated by the interstitial cell stimulating hormone (ICSH), or luteinizing hormone (LH) of the anterior pituitary (the female equivalent of ICSH).³ Testosterone is responsible for the development of secondary sex characteristics, such as the accessory sex organs, the prostate, seminal vesicles and the growth of facial, pubic and auxiliary hair. Testosterone measurements have been very helpful in evaluating hypogonadal states. Increased testosterone levels in males can be found in complete androgen resistance (testicular feminization). Common causes of decreased testosterone levels in males include: hypogonadism, orchidectomy, estrogen therapy, Klinefelter's syndrome, hypopituitarism, and hepatic cirrhosis.^{2,4}


In the female, testosterone levels are normally found to be much lower than those encountered in the healthy male. Testosterone in the female comes from three sources. It is secreted in small quantities by both the adrenal glands and the ovaries, and in healthy women 50-60% of the daily testosterone production arises from peripheral metabolism of prohormone, chiefly androstenedione. Common causes of increased serum testosterone levels in females include polycystic ovaries (Stein-Leventhal syndrome), ovarian tumors, adrenal tumors and adrenal hyperplasia. Virilization in women is associated with the administration of androgens and endogenous overproduction of testosterone. There appears to be a correlation between serum testosterone levels and the degree of virilization in women, although approximately 25% of women with varying degrees of virilism have serum testosterone levels that fall within the female reference range.

3.0 PRINCIPLE

Competitive Enzyme Immunoassay (TYPE 7):


The essential reagents required for an enzyme immunoassay include antibody, enzyme-antigen conjugate and native antigen.

Upon mixing biotinylated antibody, enzyme-antigen conjugate and a serum containing the native antigen, a competition reaction results between the native antigen and the enzyme-antigen conjugate for a limited number of antibody binding sites. The interaction is illustrated by the following equation:



Ab_{BtN} = Biotinylated Antibody (Constant Quantity)
 Ag = Native Antigen (Variable Quantity)
 EnzAg = Enzyme-antigen Conjugate (Constant Quantity)
 AgAb_{BtN} = Antigen-Antibody Complex
 EnzAgAb_{BtN} = Enzyme-antigen Conjugate -Antibody Complex
 k_a = Rate Constant of Association
 k_{-a} = Rate Constant of Disassociation
 K = k_a / k_{-a} = Equilibrium Constant

A simultaneous reaction between the biotin attached to the antibody and the streptavidin immobilized on the microwell occurs. This effects the separation of the antibody bound fraction after decantation or aspiration.



The enzyme activity in the antibody bound fraction is inversely proportional to the native antigen concentration. By utilizing several different serum references of known antigen concentration, a dose response curve can be generated from which the antigen concentration of an unknown can be ascertained.

4.0 REAGENTS

Materials Provided:

A. Testosterone Calibrators – 1ml/vial – Icons A-G

Seven (7) vials of serum reference for Testosterone at concentrations of 0 (A), 0.1 (B), 0.5 (C), 1.0 (D), 2.5 (E), 5.0 (F) and 12.0 (G) in ng/ml. Store at 2-8°C. A preservative has been added. The calibrators can be expressed in molar concentrations (nM/L) by multiplying by 3.47. For example: 1ng/ml x 3.47 = 3.47 nM/L

B. Testosterone Enzyme Reagent – 6.0 ml/vial – Icon E

One (1) ready to use vial of Testosterone (Analog)-horse radish peroxidase (HRP) conjugate in a protein stabilizing matrix with buffer, red dye, preservative, and binding protein inhibitors. Store 2-8°C.

C. Testosterone Biotin Reagent – 6.0 ml – Icon V

One (1) vial containing anti-Testosterone biotinylated purified rabbit IgG conjugate in buffer, dye and preservative. Store at 2-8°C.

D. Streptavidin Coated Plate – 96 wells – Icon J

One 96-well microplate coated with 1.0 µg/ml streptavidin and packaged in an aluminum bag with a drying agent. Store at 2-8°C.

E. Wash Solution Concentrate – 20ml/vial – Icon L

One (1) vial containing a surfactant in buffered saline. A preservative has been added. Store at 2-8°C.

F. Substrate A – 7ml/vial – Icon S^A

One (1) vial containing tetramethylbenzidine (TMB) in buffer. Store at 2-8°C.

G. Substrate B – 7ml/vial – Icon S^B

One (1) vial containing hydrogen peroxide (H₂O₂) in buffer. Store at 2-8°C.

H. Stop Solution – 8ml/vial – Icon STOP

One (1) vial containing a strong acid (1N HCl). Store at 2-8°C.

I. Product Instructions.

Note 1: Do not use reagents beyond the kit expiration date.

Note 2: Avoid extended exposure to heat and light. **Opened reagents are stable for sixty (60) days when stored at 2-8°C. Kit and component stability are identified on the label.**

Note 3: Above reagents are for a single 96-well microplate.

4.1 Required But Not Provided:

- Pipette capable of delivering 0.010ml (10µl), 0.050ml(50µl), 0.100ml (100µl) volumes with a precision of better than 1.5%.

- Dispenser(s) for repetitive deliveries of 0.50ml (50µl) .0.100ml (100 µl) and 0.350ml (350µl) volumes with a precision of better than 1.5%.
- Microplate washer or a squeeze bottle (optional).
- Microplate Reader with 450nm and 620nm wavelength absorbance capability.
- Absorbent Paper for blotting the microplate wells.
- Plastic wrap or microplate covers for incubation steps.
- Vacuum aspirator (optional) for wash steps.
- Timer.
- Quality control materials.

5.0 PRECAUTIONS

For In Vitro Diagnostic Use

Not for Internal or External Use in Humans or Animals

All products that contain human serum have been found to be non-reactive for Hepatitis B Surface Antigen, HIV 1&2 and HCV Antibodies by FDA required tests. Since no known test can offer complete assurance that infectious agents are absent, all human serum products should be handled as potentially hazardous and capable of transmitting disease. Good laboratory procedures for handling blood products can be found in the Center for Disease Control / National Institute of Health, "Biosafety in Microbiological and Biomedical Laboratories," 2nd Edition, 1988, HHS Publication No. (CDC) 88-8395.

Safe Disposal of kit components must be according to local regulatory and statutory requirement.

6.0 SPECIMEN COLLECTION AND PREPARATION

The specimens shall be blood; serum or plasma in type and the usual precautions in the collection of venipuncture samples should be observed. For accurate comparison to established normal values, a fasting morning serum sample should be obtained. The blood should be collected in a plain redtop venipuncture tube or (for plasma) in evacuated tube(s) containing heparin. Allow the blood to clot for serum samples. Centrifuge the specimen to separate the serum or plasma from the cells.

In patients receiving therapy with high biotin doses (i.e. >5mg/day), no sample should be taken until at least 8 hours after the last biotin administration, preferably overnight to ensure fasting sample.

Samples may be refrigerated at 2-8°C for a maximum period of five (5) days. If the specimen(s) cannot be assayed within this time, the sample(s) may be stored at temperatures of -20°C for up to 30 days. Avoid use of contaminated devices. Avoid repetitive freezing and thawing. When assayed in duplicate, 0.020ml (20µl) of the specimen is required.

7.0 QUALITY CONTROL

Each laboratory should assay controls at levels in the low, normal and high range for monitoring assay performance. These controls should be treated as unknowns and values determined in every test procedure performed. Quality control charts should be maintained to follow the performance of the supplied reagents. Pertinent statistical methods should be employed to ascertain trends. The individual laboratory should set acceptable assay performance limits. In addition, maximum absorbance should be consistent with past experience. Significant deviation from established performance can indicate unnoticed change in experimental conditions or degradation of kit reagents. Fresh reagents should be used to determine the reason for the variations.

8.0 REAGENT PREPARATION

- Wash Buffer**
Dilute contents of wash solution concentrate to 1000ml with distilled or deionized water in a suitable storage container. Diluted buffer can be stored at 2-30°C for up to 60 days.
- Working Substrate Solution - Stable for 1 year.**
Pour the contents of the amber vial labeled Solution 'A' into the clear vial labeled Solution 'B'. Place the yellow cap on the clear vial for easy identification. Mix and label accordingly. Store at 2 - 8°C.

Note 1: Do not use the working substrate if it looks blue.
Note 2: Do not use reagents that are contaminated or have bacteria growth.

9.0 TEST PROCEDURE

Before proceeding with the assay, bring all reagents, serum calibrators and controls to room temperature (20 - 27°C).

****Test Procedure should be performed by a skilled individual or trained professional****

- Format the microplates' wells for each serum reference, control and patient specimen to be assayed in duplicate. **Replace any unused microwell strips back into the aluminum bag, seal and store at 2-8°C.**
- Pipette 0.010 ml (10µL) of the appropriate serum reference, control or specimen into the assigned well.
- Add 0.050 ml (50µl) of the ready to use Testosterone Enzyme Reagent to all wells.
- Swirl the microplate gently for 20-30 seconds to mix.
- Add 0.050 ml (50µl) of Testosterone Biotin Reagent to all wells.
- Swirl the microplate gently for 20-30 seconds to mix.
- Cover and incubate for 60 minutes at room temperature.
- Discard the contents of the microplate by decantation or aspiration. If decanting, blot the plate dry with absorbent paper.
- Add 0.350ml (350µl) of wash buffer (see Reagent Preparation Section), decant (tap and blot) or aspirate. Repeat two (2) additional times for a total of three (3) washes. **An automatic or manual plate washer can be used. Follow the manufacturer's instruction for proper usage. If a squeeze bottle is employed, fill each well by depressing the container (avoiding air bubbles) to dispense the wash. Decant the wash and repeat two (2) additional times.**
- Add 0.100 ml (100µl) of working substrate solution to all wells (see Reagent Preparation Section). **Always add reagents in the same order to minimize reaction time differences between wells.**
DO NOT SHAKE THE PLATE AFTER SUBSTRATE ADDITION
- Incubate at room temperature for fifteen (15) minutes.
- Add 0.050ml (50µl) of stop solution to each well and gently mix for 15-20 seconds. **Always add reagents in the same order to minimize reaction time differences between wells.**
- Read the absorbance in each well at 450nm (using a reference wavelength of 620-630nm to minimize well imperfections) in a microplate reader. **The results should be read within thirty (30) minutes of adding the stop solution.**

Note: Dilute the samples suspected of concentrations higher than 12 ng/ml 1:5 and 1:10 with Testosterone 'O' ng/ml calibrator or female patient sera with a known low value for testosterone.

10.0 CALCULATION OF RESULTS

A dose response curve is used to ascertain the concentration of Testosterone in unknown specimens.

- Record the absorbance obtained from the printout of the microplate reader as outlined in Example 1.
- Plot the absorbance for each duplicate serum reference versus the corresponding Testosterone concentration in ng/ml on linear graph paper (do not average the duplicates of the serum references before plotting).
- Connect the points with a best-fit curve.
- To determine the concentration of Testosterone for an unknown, locate the average absorbance of the duplicates for each unknown on the vertical axis of the graph, find the intersecting point on the curve, and read the concentration (in ng/ml) from the horizontal axis of the graph (the duplicates of the unknown may be averaged as indicated). In the following example, the average absorbance (1.764) intersects the dose response curve at (0.57ng/ml) Testosterone concentration (See Figure 1).


Note: Computer data reduction software designed for ELISA assays may also be used for the data reduction. **If such software is utilized, the validation of the software should be ascertained.**

EXAMPLE 1

Sample I.D.	Well Number	Abs (A)	Mean Abs (B)	Value (ng/ml)
Cal A	A1	2.780	2.787	0
	B1	2.794		
Cal B	C1	2.576	2.611	0.1
	D1	2.646		
Cal C	E1	1.789	1.877	0.5
	F1	1.965		
Cal D	G1	1.391	1.392	1.0
	H1	1.393		
Cal E	A2	0.780	0.788	2.5
	B2	0.796		
Cal F	C2	0.530	0.538	5.0
	D2	0.547		
Cal G	E2	0.301	0.308	12.0
	F2	0.314		
Ctrl 1	G2	1.040	0.760	1.61
	H2	1.045		
Patient	A3	1.751	1.764	0.57
	B3	1.778		

*The data presented in Example 1 and Figure 1 is for illustration only and **should not** be used in lieu of a standard curve prepared with each assay.

Figure 1



11.0 Q.C. PARAMETERS

In order for the assay results to be considered valid the following criteria should be met:

- The absorbance (OD) of calibrator 0 ng/ml should be ≥ 1.3
- Four out of six quality control pools should be within the established ranges.

12.0 RISK ANALYSIS

The MSDS and Risk Analysis Form for this product are available on request from Monobind Inc.

12.1 Assay Performance

- It is important that the time of reaction in each well is held constant to achieve reproducible results.
- Pipetting of samples should not extend beyond ten (10) minutes to avoid assay drift.
- Highly lipemic, hemolyzed or grossly contaminated specimen(s) should not be used.
- If more than one (1) plate is used, it is recommended to repeat the dose response curve.
- The addition of substrate solution initiates a kinetic reaction, which is terminated by the addition of the stop solution. Therefore, the substrate and stop solution should be added in the same sequence to eliminate any time-deviation during reaction.
- Plate readers measure vertically. Do not touch the bottom of the wells.
- Failure to remove adhering solution adequately in the aspiration or decantation wash step(s) may result in poor replication and spurious results.
- Use components from the same lot. No intermixing of reagents from different batches.
- Accurate and precise pipetting, as well as following the exact time and temperature requirements prescribed are essential.

Any deviation from Monobind's IFU may yield inaccurate results.

- All applicable national standards, regulations and laws, including, but not limited to, good laboratory procedures, must be strictly followed to ensure compliance and proper device usage.
- It is important to calibrate all the equipment e.g. Pipettes, Readers, Washers and/or the automated instruments used with this device, and to perform routine preventative maintenance.
- Risk Analysis: as required by CE Mark IVD Directive 98/79/EC - for this and other devices, made by Monobind, can be requested via email from Monobind@monobind.com.

12.2 Interpretation

- Measurements and interpretation of results must be performed by a skilled individual or trained professional.** Laboratory results alone are only one aspect for determining patient care and should not be the sole basis for therapy, particularly if the results conflict with other determinants.
- The reagents for the procedure have been formulated to eliminate maximal interference; however, potential interaction between rare serum specimens and test reagents can cause erroneous results. Heterophilic antibodies often cause these interactions and have been known to be problems for all kinds of immunoassays. (*Boscato LM, Stuart MC, 'Heterophilic antibodies: a problem for all immunoassays' Clin. Chem 1988:3427-33*). For diagnostic purposes the results from this assay should be used in combination with clinical examination, patient's history and, all other clinical findings.
- For valid test results, adequate controls and other parameters must be within the listed ranges and assay requirements.
- If test kits are altered, such as by mixing parts of different kits, which could produce false test results, or if results are incorrectly interpreted, **Monobind shall have no liability.**
- If computer controlled data reduction is used to interpret the results of the test, it is imperative that the predicted values for the calibrators fall within 10% of the assigned concentrations.

13.0 EXPECTED RANGES OF VALUES

In agreement with established reference intervals⁵ for a "normal" adult population, the expected ranges for the Testosterone AccuBind® ELISA Test System are detailed in Table 1.

TABLE 1
Expected Values for Testosterone EIA Test System (ng/ml)

Boys Before Puberty	0.1 – 3.7
Male	2.5 – 10.0
Female	0.2 – 0.95

It is important to keep in mind that establishment of a range of values, which can be expected to be found by a given method for a population of "normal" persons, is dependent upon a multiplicity of factors: the specificity of the method, the population tested and the precision of the method in the hands of the analyst. For these reasons, each laboratory should depend upon the range of expected values established by the Manufacturer only until an in-house range can be determined by the analysts using the method with a population indigenous to the area in which the laboratory is located.

14.0 PERFORMANCE CHARACTERISTICS

14.1 Precision

The within and between assay precision of the Testosterone AccuBind® ELISA Test System were determined by analyses on three different levels of pool control sera. The number, mean values, standard deviation and coefficient of variation for each of these control sera are presented in Table 2 and Table 3.

TABLE 2
Within Assay Precision (Values in ng/ml)

Sample	N	X	σ	C.V.
Low	22	1.63	0.16	9.8%
Normal	22	9.14	0.44	4.8%
High	22	14.22	0.79	5.6%

TABLE 3
Between Assay Precision (Values in ng/ml)

Sample	N	X	σ	C.V.
Low	24	1.72	0.16	9.1%
Normal	24	7.06	0.69	9.7%
High	24	13.08	1.03	7.9%

*As measured in several experiments in duplicate over a ten day period.

14.2 Sensitivity

The Testosterone AccuBind® ELISA Test System has a sensitivity of 0.576 pg. This is equivalent to a sample containing a concentration of 0.0576 ng/ml. The sensitivity was ascertained by determining the variability of the 0 ng/ml serum calibrator and using the 2σ (95% certainty) statistic to calculate the minimum dose.

14.3 Accuracy

The Testosterone AccuBind® ELISA Test System was compared with a chemiluminescence immunoassay method. Biological specimens from low, normal and high Testosterone level populations were used. The values ranged from 0.29 ng/ml – 21.9ng/ml. The total number of such specimens was 58. The least square regression equation and the correlation coefficient were computed for this Testosterone EIA in comparison with the reference method. The data obtained is displayed in Table 4.

TABLE 4

Method	Mean (x)	Least Square Regression Analysis	Correlation Coefficient
Monobind (y)	3.12	y = -0.265 + 0.944(x)	0.985
Reference (X)	3.02		

Only slight amounts of bias between this method and the reference method are indicated by the closeness of the mean values. The least square regression equation and correlation coefficient indicates excellent method agreement.

14.4 Specificity

The % cross reactivity of the testosterone antibody to selected substances was evaluated by adding the interfering substance to a serum matrix at various concentrations. The cross-reactivity was calculated by deriving a ratio between dose of interfering substance to dose of Testosterone needed to displace the same amount of labeled analog.

Substance	Cross Reactivity
Testosterone	1.0000
Androstenedione	0.0009
Dihydrotestosterone	0.0178
Cortisone	<0.0001
Corticosterone	<0.0001
Cortisol	<0.0001
Spirolactone	<0.0001
Progesterone	<0.0001
17α-OH Progesterone	<0.0001
DHEA sulfate	<0.0001
Estradiol	<0.0001
Estrone	<0.0001
Estriol	<0.0001
Hemolysis	<0.0001
Rubella	<0.0001
Lipemia	<0.0001

15.0 REFERENCES

- Dorfman, RI and Shipley, RA, ED: Androgens, New York.: John Wiley and Sons, 1956.
- Horton R, Tait JF: Androstenedione production and interconversion rates measured in peripheral blood and studies on the possible site of conversion to testosterone. J.Clin Invest 45: 301-303, 1966.
- Faiman C and Winter, JSD, Reyes, FI, Clin *Obstet Gynaecol*, 3, 467 (1976).
- Sizonenka, PC, *Pediatrician*, 14, 191 (1987).
- Cummings DC, Wall SR: Non sex hormone binding globulin bound testosterone as a marker for hyperandrogenism. J. Clin Endocrinol Metab. 61:873-876, 1985.
- Lashansky, G, et. al., *J Clin Endocrinol Metab*, 58, 674 (1991)
- Tietz, NW, ED: Clinical Guide to Laboratory Tests, 3rd ed. Philadelphia, WA Saunders Co, 1995.

Size	96(A)	192(B)
Reagent (fill)	A) 1ml set	1ml set
	B) 1 (6ml)	2 (6ml)
	C) 1 (6ml)	2 (6ml)
	D) 1 plate	2 plates
	E) 1 (20ml)	1 (20ml)
	F) 1 (7ml)	2 (7ml)
	G) 1 (7ml)	2 (7ml)
	H) 1 (8ml)	2 (8ml)

For Orders and Inquires, please contact



Tel: +1 949.951.2665 Mail: info@monobind.com
Fax: +1 949.951.3539 Fax: www.monobind.com



CEpartner4U, Esdoornlaan 13
3951 DBMaarn, The Netherlands
www.cepartner4u.eu

Please visit our website to learn more about our products and services.

Glossary of Symbols
(EN 980/ISO 15223)

In Vitro - Diagnostic Medical Device	Temperature Limitation Storage Condition (2-8° C)	Consult Instructions for Use
Catalogue Number	Contains Sufficient Test for Σ	Batch Code
Used By (Expiration Day)	Date of Manufacturer	Manufacturer
Authorized Rep in European Country	European Conformity	