

**VALUE ASSIGNMENT FOR BIO-RAD**  
**MASTER LOT 5466(1-3)      Exp date:2021-Aug-31**

AccuLite® CLIA	54661		54662		54663	
	Mean	Range	Mean	Range	Mean	Range
<b>Tumor Markers</b>						
Ferritin in ng/ml	53.32	35.73-70.92	365.84	245.11-486.56	799.99	535.99->800
CEA Next Generation in ng/ml	1.57	1.05-2.09	19.95	13.37-26.54	70.63	47.32-93.94
CEA in ng/ml	1.84	1.23-2.45	23.07	15.46-30.69	83.63	56.03-111.23
AFP in ng/ml	7.67	5.14-10.20	67.87	46.47-90.26	218.22	146.21-290.23
PSA in ng/ml	<0.5		5.68	3.81-7.56	23.68	15.87-31.50
fPSA in ng/ml	<0.5		2.58	1.73-3.42	>10	
CA-125 in U/ml	16.16	10.83-21.49	51.02	34.19-67.86	176.39	118.18-234.60
CA-19-9 in U/ml	17.40	11.66-23.14	31.93	21.39-42.46	82.98	55.59-110.36
CA 15-3 in U/ml	24.68	16.54-32.83	47.32	31.70-62.93	80.03	53.62-106.44
PRL in ng/ml	11.90	7.97-15.82	46.48	31.14-61.82	>100	
HCG-XR in mIU/ml	3.09	2.07-4.10	43.98	29.46-58.49	225.68	151.20-300.15
HCG in mIU/ml	4.36	2.92-5.80	71.21	47.71-94.72	>200	
TG in ng/ml	5.64	3.78-7.51	35.89	24.05-47.73	83.56	55.99-111.14

DOCUMENT HISTORY

PREPARED BY: *MSW* DEPT: QC  
 APPROVED BY: *Frederick* DEPT: Administration  
 REVISION: 0

VERIFIED BY: *Ashatola* DEPT: QA  
 EFFECTIVE DATE: 2018-Aug-27  
 DCO: N/A