



VALUE ASSIGNMENT FOR BIO-RAD

MASTER LOT 2988(1-3)

Exp: 2021-May-31

AccuLite® CLIA	29881		29882		29883	
	Mean	Range	Mean	Range	Mean	Range
Cardiac Markers						
CKMB in ng/ml	2.06	1.38 - 2.74	12.86	8.62 - 17.1	80.49	53.93 - 107.05
hsCRP in µg/ml	0.65	0.44 - 0.86	2.24	1.50 - 2.98	7.35	4.93 - 9.77
Myoglobin in ng/ml	66.27	44.4 - 88.14	175.82	117.8 - 233.84	363.79	243.74 - >400

DOCUMENT HISTORY

PREPARED BY: NSW DEPT: QC VERIFIED BY: Ashatola DEPT: QA
 APPROVED BY: Fred Jones DEPT: Administration EFFECTIVE DATE: 2018-Feb-14
 REVISION: 0 DCO: 0