



**VALUE ASSIGNMENT FOR BIO-RAD**
**MASTER LOT 5401(1-3) Exp date: 2020-Jun-30**

Acculite® CLIA	54011		54012		54013	
	Mean	Range	Mean	Range	Mean	Range
<b>Liquichek Maternal Serum 1st Trimester</b>						
hCG in mIU/mL	5550.0	3718.5 - 7381.5	20200.0	13534.0 - 26866.0	47000.0	31490.0 - 62510.0
hCG-XR in mIU/mL	3956.71	2651.00 - 5262.43	18920.83	12676.96 - 25164.71	40169.71	26913.71 - 53425.72
PAPP-A in µg/ml	3.95	2.64 - 5.25	13.71	9.18 - 18.23	34.45	23.08 - 45.82
fβ-hCG-XR in ng/mL	16.12	10.80 - 21.44	72.50	48.57 - 96.42	150.82	101.05 - 200.59

**Note 1:** The hCG and hCG-XR data was obtained by diluting the sample on a 1:1000 and 1:100 dilution respectively. The dilution factor was later applied to generate the correct value for each level.

**Note 2:** PAPP-A data is reported in µg/ml. To convert µg/ml into mIU/L, multiply the values by 156.25.  
Ex. 32 µg/ml x 156.25 = 5000 mIU/L

## DOCUMENT HISTORY

PREPARED BY:  DEPT: QC      VERIFIED BY: A. Shatola DEPT: QA  
 APPROVED BY:  DEPT: Administration      EFFECTIVE DATE: 2018-Sep-14  
 REVISION: 0      DCO: N/A