




**VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 4226(1-3) EXP: 2021-SEP-30**

| AccuLite® CLIA | 42261 | | 42262 | | 42263 | |
|-----------------------|--------|-----------------|--------|-----------------|----------|---------------------|
| | Mean | Range | Mean | Range | Mean | Range |
| Fertility | | | | | | |
| Ferritin in ng/ml | 13.50 | 9.05 - 17.96 | 52.90 | 35.44 - 70.35 | 101.73 | 68.16 - 135.30 |
| AFP in ng/ml | 22.39 | 15.00 - 29.77 | 80.00 | 53.60 - 106.40 | 316.10 | 211.79 - 420.41 |
| LH in mIU/ml | 135.15 | 90.55 - 179.75 | 54.24 | 36.34 - 72.14 | 32.49 | 21.77 - 43.21 |
| hCG in mIU/ml | <2 | N/A | 32.68 | 21.90 - 43.47 | 24570.67 | 16462.35 - 32678.99 |
| Testosterone in ng/ml | 0.33 | 0.22 - 0.44 | 3.47 | 2.32 - 4.61 | 8.88 | 5.95 - 11.80 |
| Progesterone in ng/ml | 25.48 | 17.07 - 33.88 | 44.72 | 29.96 - 59.48 | >60 | N/A |
| E2 in pg/ml | 150.97 | 101.15 - 200.79 | 190.28 | 127.49 - 253.08 | 825.08 | 552.80 - 1097.35 |

DOCUMENT HISTORY

PREPARED BY:  DEPT: QC
 VERIFIED BY:  DEPT: QA
 APPROVED BY:  DEPT: Administration
 EFFECTIVE DATE: 2018-Dec-17
 REVISION: 0
 DCO: N/A