

VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 2398(1-3)
Exp: 2021-Mar-31

AccuBind® ELISA	23981		23982		23983	
	Mean	Range	Mean	Range	Mean	Range
Tumor Markers						
Ferritin in ng/ml	26.92	18.03 - 35.80	383.72	257.09 - 510.35	> 600	
CEA in ng/ml	4.41	2.96 - 5.87	37.24	24.95 - 49.52	127.03	85.11 - 168.95
AFP in ng/ml	7.81	5.24 - 10.39	93.66	62.75 - 124.56	292.19	195.77 - 388.62
PSA in ng/ml	< 0.5		9.02	6.04 - 11.99	35.69	23.91 - 47.47
fPSA in ng/ml	<0.3		4.19	2.80 - 5.57	> 10	
CA-125 in U/ml	27.73	18.58 - 36.88	77.15	51.69 - 102.61	203.51	136.35 - 270.67
CA-19-9 in U/ml	6.67	4.47 - 8.87	23.78	15.94 - 31.63	78.71	52.74 - 104.68
CEA Next Generation in ng/ml	3.34	2.23 - 4.44	33.35	22.35 - 44.36	96.10	64.39 - 127.82
PRL in ng/ml	11.68	7.83 - 15.53	36.52	24.47 - 48.57	> 100	
HCG in mIU/ml	3.27	2.19 - 4.34	50.04	33.53 - 66.56	240.24	160.96 - > 250
TG in ng/ml	5.95	3.99 - 7.91	55.84	37.42 - 74.27	137.86	92.37 - 183.36
CA 15-3 in U/ml	18.25	12.23 - 24.27	39.62	26.54 - 52.69	74.63	50.00 - 99.26
HCG-XR in mIU/ml	3.24	2.17 - 4.30	45.46	30.46 - 60.46	201.21	134.81 - 267.61

DOCUMENT HISTORY

PREPARED BY: *MSW* DEPT: QC
 APPROVED BY: *Fred Lopez* DEPT: Administration
 REVISION: 0

VERIFIED BY: *A Shatola* DEPT: QA
 EFFECTIVE DATE: 2019-Jan-18
 DCO: N/A

