

VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 2398(1-3)
Exp: 2021-Mar-31

| AccuLite® CLIA | 23981 | | 23982 | | 23983 | |
|------------------------------|-------|---------------|--------|-----------------|--------|-----------------|
| | Mean | Range | Mean | Range | Mean | Range |
| Tumor Markers | | | | | | |
| Ferritin in ng/ml | 30.16 | 20.20 - 40.11 | 392.18 | 262.76 - 521.60 | > 700 | |
| CEA in ng/ml | 3.96 | 2.66 - 5.27 | 36.03 | 24.14 - 47.92 | 117.75 | 78.89 - 156.61 |
| AFP in ng/ml | 7.34 | 4.92 - 9.76 | 97.00 | 64.99 - 129.01 | 276.79 | 185.45 - 368.13 |
| PSA in ng/ml | < 1 | | 8.31 | 5.57 - 11.05 | 35.35 | 23.68 - 47.01 |
| fPSA in ng/ml | < 0.3 | | 4.21 | 2.82 - 5.60 | > 10 | |
| CA-125 in U/ml | 25.51 | 17.09 - 33.93 | 72.79 | 48.77 - 96.81 | 207.98 | 139.35 - 276.61 |
| CA-19-9 in U/ml | 5.56 | 3.73 - 7.40 | 20.68 | 13.85 - 27.50 | 69.65 | 46.66 - 92.63 |
| CEA Next Generation in ng/ml | 2.75 | 1.84 - 3.66 | 29.40 | 19.70 - 39.11 | 97.39 | 65.25 - 129.53 |
| PRL in ng/ml | 11.65 | 7.80 - 15.49 | 37.05 | 24.82 - 49.28 | > 100 | |
| HCG in mIU/ml | 2.85 | 1.91 - 3.79 | 48.79 | 32.69 - 64.90 | 245.61 | 164.56 - > 250 |
| TG in ng/ml | 5.27 | 3.53 - 7.01 | 51.41 | 34.45 - 68.38 | 142.80 | 95.68 - 189.93 |
| CA 15-3 in U/ml | 20.29 | 13.60 - 26.99 | 43.26 | 28.99 - 57.54 | 78.39 | 52.52 - 104.26 |
| HCG-XR in mIU/ml | 3.59 | 2.41 - 4.78 | 47.91 | 32.10 - 63.72 | 237.87 | 159.37 - 316.36 |

DOCUMENT HISTORY

PREPARED BY: MSW DEPT: QC
 APPROVED BY: Fred Lopez DEPT: Administration
 REVISION: 0

VERIFIED BY: AShatola DEPT: QA
 EFFECTIVE DATE: 2019-Jan-18
 DCO: N/A