

VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 2397(1-3) Exp: 2020-Nov-30

AccuLite® CLIA	23971		23972		23973	
	Mean	Range	Mean	Range	Mean	Range
Tumor Markers						
Ferritin in ng/ml	26.17	17.54-34.81	282.74	189.44-376.05	764.14	511.98->700
CEA in ng/ml	2.05	1.37-2.73	37.38	25.04-49.71	75.20	50.38-100.02
AFP in ng/ml	8.42	5.64-11.19	85.17	57.06-113.27	268.75	224.56-312.94
PSA in ng/ml	<1		8.23	5.51-10.94	34	22.78-45.22
fPSA in ng/ml	<0.3		3.24	2.17-4.31	>10	
CA-125 in U/ml	21.70	17.59-25.80	63.10	56.22-69.98	204.08	136.73-271.43
CA-19-9 in U/ml	5.46	3.65-7.26	21	14.07-27.92	72.08	48.29-95.86
CEA Next Generation in ng/ml	0.96	0.64-1.28	26.77	17.94-35.60	100.74	67.50-133.99
PRL in ng/ml	11.63	7.79-15.47	39.43	26.42-52.45	>100	
HCG in mIU/ml	2.89	1.92-3.86	44.59	29.88-59.31	227.17	152.20-302.13
TG in ng/ml	5.25	3.52-6.98	47.22	31.64-62.80	147.91	99.10-196.72
CA 15-3 in U/ml	18.16	12.16-24.15	37.18	24.91-49.45	62.43	41.83-83.03

DOCUMENT HISTORY

PREPARED BY: *NSW* DEPT: QC VERIFIED BY: *Ashatola* DEPT: QA
 APPROVED BY: *Fred Lopez* DEPT: Administration EFFECTIVE DATE: 2018-Oct-04
 REVISION: 0 DCO: N/A