

VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 2399(1-3) Exp: 2021-Aug-31

AccuLite® CLIA	23991		23992		23993	
	Mean	Range	Mean	Range	Mean	Range
Tumor Markers						
Ferritin in ng/ml	31.72	21.25 - 42.19	390.62	261.72 - 519.53	> 700	
CEA in ng/ml	3.66	2.45 - 4.86	41.05	27.51 - 54.60	124.17	83.19 - 165.14
AFP in ng/ml	7.61	5.10 - 10.12	103.06	69.05 - 137.07	297.82	199.54 - 396.10
PSA in ng/ml	< 1		8.87	5.95 - 11.80	36.70	24.59 - 48.82
fPSA in ng/ml	< 0.3		4.07	2.73 - 5.41	> 10	
CA-125 in U/ml	28.17	18.88 - 37.47	87.48	58.61 - 116.35	264.63	177.30 - 351.95
CA-19-9 in U/ml	5.91	3.96 - 7.86	23.47	15.73 - 31.22	73.01	48.92 - 97.10
CEA Next Generation in ng/ml	2.52	1.69 - 3.35	31.88	21.36 - 42.40	97.41	65.27 - 129.56
PRL in ng/ml	11.04	7.39 - 14.68	35.65	23.88 - 47.41	> 100	
HCG in mIU/ml	2.97	1.99 - 3.95	53.60	35.91 - 71.29	238.59	159.85 - >250
TG in ng/ml	5.53	3.71 - 7.36	53.00	35.51 - 70.49	143.38	96.09 - 190.69
CA 15-3 in U/ml	18.83	12.62 - 25.05	46.23	30.98 - 61.49	99.88	66.92 - 132.84
HCG-XR in mIU/ml	2.02	1.36 - 2.69	45.65	30.58 - 60.71	208.64	139.79 - 277.50

DOCUMENT HISTORY

PREPARED BY: *NSP* DEPT: QC VERIFIED BY: *Ashatola* DEPT: QA
 APPROVED BY: *Fred Lopez* DEPT: Administration EFFECTIVE DATE: 2019-May-20
 REVISION: 0 DCO: N/A