

VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT 2399(1-3)			Exp: 2021-Aug-31			
AccuBind® ELISA	23991		23992		23993	
	Mean	Range	Mean	Range	Mean	Range
Tumor Markers						
Ferritin in ng/ml	27.58	18.48 - 36.68	362.34	242.76 - 481.91	>700	
CEA in ng/ml	4.39	2.94 - 5.83	40.82	27.35 - 54.29	132.36	88.68 - 176.03
AFP in ng/ml	8.19	5.49 - 10.89	106.55	71.39 - 141.71	293.08	196.36 - 389.79
PSA in ng/ml	< 0.5		8.92	5.97 - 11.86	34.87	23.36 - 46.38
fPSA in ng/ml	<0.3		4.19	2.81 - 5.57	> 10	
CA-125 in U/ml	31.60	21.17 - 42.02	92.21	61.78 - 122.64	278.34	186.49 - 370.19
CA-19-9 in U/ml	6.23	4.18 - 8.29	24.04	16.11 - 31.98	83.13	55.70 - 110.56
CEA Next Generation in ng/ml	3.08	2.06 - 4.09	32.55	21.81 - 43.30	96.24	64.48 - 128.00
PRL in ng/ml	11.17	7.48 - 14.85	35.67	23.90 - 47.44	> 100	
HCG in mIU/ml	3.29	2.21 - 4.38	51.09	34.23 - 67.95	231.85	155.34 - >250
TG in ng/ml	6.33	4.24 - 8.41	54.37	36.43 - 72.31	147.27	98.67 - 195.87
CA 15-3 in U/ml	18.29	12.25 - 24.32	41.12	27.55 - 54.69	90.99	60.97 - 121.02
HCG-XR in mIU/ml	2.75	1.84 - 3.65	47.08	31.55 - 62.62	197.92	132.60 - 263.23

DOCUMENT HISTORY			
PREPARED BY: <u>MSW</u>	DEPT: QC	VERIFIED BY: <u>Ashatola</u>	DEPT: QA
APPROVED BY: <u>[Signature]</u>	DEPT: Administration	EFFECTIVE DATE: 2019-May-20	
REVISION: 0		DCO: N/A	

