

VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 5468(1-3) Exp date:2022-August-31

AccuBind® ELISA	54681		54682		54683	
	Mean	Range	Mean	Range	Mean	Range
Tumor Markers						
Ferritin in ng/ml	56.67	37.97-75.37	255.12	170.93-339.31	574.32	384.79-763.84
CEA Next Generation in ng/ml	2.92	1.96-3.88	26.98	18.08-35.88	86.47	57.93-115.0
CEA in ng/ml	3.67	2.46-4.88	28.86	19.34-38.39	94.11	63.05-125.17
AFP in ng/ml	7.83	5.25-10.41	81.01	54.28-107.74	233.13	153.20-310.06
PSA in ng/ml	<0.5		7.11	4.76-9.45	28.03	18.78-37.28
fPSA in ng/ml	<0.5		3.67	2.46-4.88	>10	
CA-125 in U/ml	26.50	17.75-35.24	74.58	49.97-99.19	235.03	157.47-312.58
CA-19-9 in U/ml	16.96	11.36-22.56	34.51	23.12-45.90	89.06	59.67-118.45
CA 15-3 in U/ml	30.43	20.39-40.47	55.72	37.33-74.10	104.72	70.16-139.28
PRL in ng/ml	12.11	8.12-16.11	36.38	24.37-48.38	>100	
HCG-XR in mIU/ml	4.00	2.68-5.32	67.82	45.44-90.20	332.66	222.88-442.43
HCG in mIU/ml	5.11	3.42-6.79	68.42	45.84-91.00	>200	
TG in ng/ml	6.78	4.54-9.02	35.81	23.99-47.63	92.72	62.12-123.31

DOCUMENT HISTORY

 PREPARED BY: RSW DEPT: QC

 VERIFIED BY: Ashatola DEPT: QA

 APPROVED BY: Fred Lopez DEPT: Administration

EFFECTIVE DATE: 2019-July-30

REVISION: 0

DCO: N/A