

VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 5467(1-3) Exp date:2022-Mar-31

AccuLite® CLIA	54671		54672		54673	
	Mean	Range	Mean	Range	Mean	Range
Tumor Markers						
Ferritin in ng/ml	50.03	33.52 - 66.55	276.23	185.07 - 367.38	563.08	377.26 - >800
CEA Next Generation in ng/ml	2.91	1.95 - 3.87	30.18	20.22 - 40.14	98.95	66.29 - 131.60
CEA in ng/ml	2.57	1.72 - 3.42	27.21	18.23 - 36.19	93.97	62.96 - 124.98
AFP in ng/ml	6.88	4.61 - 9.15	81.38	54.53 - 108.24	232.99	156.10 - 309.88
PSA in ng/ml	<0.5		5.76	3.86 - 7.66	26.02	17.43 - 34.60
fPSA in ng/ml	<0.5		3.10	2.07 - 4.12	>10	
CA-125 in U/ml	20.52	13.75 - 27.30	63.05	42.25 - 83.86	196.24	131.48 - 261.01
CA-19-9 in U/ml	12.08	8.10 - 16.07	26.90	18.02 - 35.77	75.81	50.79 - 100.83
CA 15-3 in U/ml	26.99	17.69 - 36.29	53.34	35.99 - 70.69	90.95	57.78 - 124.12
PRL in ng/ml	13.52	9.06 - 17.98	39.79	26.66 - 52.92	>100	
HCG-XR in mIU/ml	3.19	2.14 - 4.25	58.01	38.87 - 77.16	259.68	173.99 - 345.38
HCG in mIU/ml	4.11	2.75 - 5.47	72.22	48.39 - 96.06	>200	
TG in ng/ml	5.19	3.48 - 6.90	29.91	20.04 - 39.78	77.91	52.20 - 103.62

DOCUMENT HISTORY

 PREPARED BY: MMO DEPT: QC

 VERIFIED BY: AShatoka DEPT: QA

 APPROVED BY: Fred Lopez DEPT: Administration

EFFECTIVE DATE: 2019-Feb-20

REVISION: 0

DCO: N/A