

**VALUE ASSIGNMENT FOR BIO-RAD**  
**MASTER LOT 5468(1-3) Exp date:2022-August-31**

AccuLite® CLIA	54681		54682		54683	
	Mean	Range	Mean	Range	Mean	Range
<b>Tumor Markers</b>						
Ferritin in ng/ml	51.09	34.23-67.95	268.65	179.99-357.30	642.05	430.17->800
CEA Next Generation in ng/ml	2.76	1.85-3.67	24.68	16.54-32.83	83.40	55.88-110.93
CEA in ng/ml	2.91	1.95-3.87	26.92	18.04-35.81	92.61	62.05-123.17
AFP in ng/ml	6.58	4.41-8.75	79.53	53.29-105.78	227.23	152.25-302.22
PSA in ng/ml	<0.5		6.77	4.53-9.0	27.53	18.44-36.61
fPSA in ng/ml	<0.5		4.44	2.97-5.90	>10	
CA-125 in U/ml	25.36	16.99-33.73	76.96	51.57-102.36	243.60	163.21-323.98
CA-19-9 in U/ml	11.41	7.64-15.17	26.23	17.57-34.88	73.85	47.20-100.51
CA 15-3 in U/ml	29.22	19.58-38.86	55.42	37.13-73.71	102.60	68.74-136.46
PRL in ng/ml	12.53	8.40-16.67	38.10	25.53-50.67	>100	
HCG-XR in mIU/ml	2.97	1.99-3.95	62.81	42.08-83.54	322.28	215.93-428.63
HCG in mIU/ml	4.33	2.90-5.76	65.97	44.20-87.74	>200	
TG in ng/ml	6.53	4.37-8.68	36.78	24.64-48.92	97.05	65.02-129.08

DOCUMENT HISTORY

PREPARED BY: *MSW* DEPT: QC  
 APPROVED BY: *Frederick* DEPT: Administration  
 REVISION: 0

VERIFIED BY: *Ashatka* DEPT: QA  
 EFFECTIVE DATE: 2019-July-20  
 DCO: N/A