



VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT6761(1-3)			Exp:2021-Oct-31			
AccuLite® CLIA	67611		67612		67613	
	Mean	Range	Mean	Range	Mean	Range
Cardiac Markers						
CKMB in ng/ml	2.87	1.93 - 3.82	13.98	9.37 - 18.60	65.87	44.13 - 87.61
hsCRP in µg/ml	0.94	0.63 - 1.25	2.22	1.49 - 2.96	6.28	4.21 - 8.35
Troponin in ng/ml	<1		0.66	0.44 - 0.87	3.54	2.37 - 4.70
Myoglobin in ng/ml	61.11	40.94 - 81.28	145.57	97.53 - 193.60	355.01	237.86 - 472.17

Note: Monobind evaluates levels 1, 2 and 3 in the Bio-Rad control.

It does not evaluate levels 1A, 1B and 1C due to minimal difference in value recognition.

DOCUMENT HISTORY

PREPARED BY: NSW DEPT: QC VERIFIED BY: A Shatola DEPT: QA
 APPROVED BY: Fred Johnson DEPT: Administration EFFECTIVE DATE: 2019-Aug-05
 REVISION: 0 DCO: N/A