

VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 5402(1-3) Exp date:2021-Apr-30

AccuBind® ELISA	54021		54022		54023	
	Mean	Range	Mean	Range	Mean	Range
Liquichek Maternal Serum 1st Trimester						
hCG in mIU/mL	3731.55	2500.14 - 4962.96	19364.35	12974.11 - 25754.58	45950.55	30786.87 - 61114.23
hCG-XR in mIU/mL	4749.63	3182.25 - 6317.01	18922.07	12677.79 - 25166.35	44690.01	29942.31 - 59437.72
PAPP-A in µg/ml	4.49	3.01 - 5.98	13.63	9.13 - 18.13	38.23	25.61 - 50.85
fβ-hCG in ng/ml	16.23	10.88 - 21.59	60.36	40.44 - 80.27	125.08	83.80 - 166.36

Note 1: The HCG and HCG-XR data was obtained by diluting the sample on a 1:1000 dilution and 1:100 dilution respectively . The dilution factor was later applied to generate the correct value for each level.

Note 2: PAPP-A data is being reported in µg/ml. To convert µg/ml into mIU/l multiply the values x 156.25

DOCUMENT HISTORY

 PREPARED BY: NSW DEPT: QC

 VERIFIED BY: A Shatola DEPT: QA

 APPROVED BY: Fred Johnson DEPT: Administration

EFFECTIVE DATE: 2019-July-26

REVISION: 0

DCO: N/A