




VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 5402(1-3) Exp date:2021-Apr-30

Acculite® CLIA	54021		54022		54023	
	Mean	Range	Mean	Range	Mean	Range
Liquichek Maternal Serum 1st Trimester						
hCG in mIU/mL	3525.00	2361.75 - 4688.25	18616.67	12473.17 - 24760.17	45241.67	30311.92 - 60171.42
hCG-XR in mIU/mL	4401.50	2949.01 - 5854.00	19313.70	12940.18 - 25687.22	42368.80	28387.10 - 56350.50
PAPP-A in µg/ml	4.48	3.00 - 5.95	12.44	7.83 - 17.05	32.49	21.77 - 43.21
fβ-hCG-XR in ng/mL	14.06	9.42 - 18.70	58.40	39.13 - 77.67	118.70	79.53 - 157.87

Note 1: The HCG and HCG-XR data was obtained by diluting the sample on a 1:1000 dilution and 1:100 dilution respectively . The dilution factor was later applied to generate the correct value for each level.

Note 2: PAPP-A data is being reported in µg/ml. To convert µg/ml into mIU/l multiply the values x 156.25

DOCUMENT HISTORY

PREPARED BY:  DEPT: QC VERIFIED BY:  DEPT: QA
 APPROVED BY:  DEPT: Administration EFFECTIVE DATE: 2019-July-26
 REVISION: 0 DCO: N/A