

VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 5467(1-3) Exp date:2022-Mar-31

AccuBind® ELISA	54671		54672		54673	
	Mean	Range	Mean	Range	Mean	Range
Tumor Markers						
Ferritin in ng/ml	53.56	35.89 - 71.24	264.52	177.23 - 351.82	559.88	375.12 - >800
CEA Next Generation in ng/ml	3.09	2.07 - 4.11	28.70	19.23 - 38.17	89.28	59.82 - 118.74
CEA in ng/ml	2.93	1.96 - 3.90	28.11	18.83 - 37.38	100.89	67.60 - 134.19
AFP in ng/ml	7.89	5.29 - 10.49	87.67	58.74 - 116.60	240.25	160.97 - 319.54
PSA in ng/ml	<0.5		5.81	3.89 - 7.72	25.66	17.19 - 34.13
fPSA in ng/ml	<0.5		3.18	2.13 - 4.23	>10	
CA-125 in U/ml	22.80	15.28 - 30.32	63.91	42.82 - 85.01	198.68	133.12 - 264.24
CA-19-9 in U/ml	18.09	12.12 - 24.07	38.08	25.51 - 50.64	86.13	57.71 - 114.56
CA 15-3 in U/ml	24.81	16.62 - 32.99	48.91	31.20 - 66.61	86.87	53.58 - 120.17
PRL in ng/ml	13.72	9.19 - 18.25	40.93	27.42 - 54.44	>100	
HCG-XR in mIU/ml	4.13	2.77 - 5.49	59.63	39.95 - 79.31	266.45	178.52 - 354.38
HCG in mIU/ml	4.56	3.05 - 6.06	70.46	47.21 - 93.71	>250	
TG in ng/ml	5.56	3.72 - 7.39	28.74	19.26 - 38.22	67.22	45.03 - 89.40

DOCUMENT HISTORY

 PREPARED BY: NSW DEPT: QC

 VERIFIED BY: Ashatola DEPT: QA

 APPROVED BY: Fred Jones DEPT: Administration

EFFECTIVE DATE: 2019-Feb-20

REVISION: 0

DCO: N/A