



VALUE ASSIGNMENT FOR BIO-RAD

MASTER LOT 8781(1-3)

Exp: 2022-09-30

AccuLite® CLIA	87811		87812		87813	
	Mean	Range	Mean	Range	Mean	Range
Cardiac Markers						
CKMB in ng/ml	1.15	0.77-1.53	12	8.04-15.96	65.46	43.86-87.07
hsCRP in µg/ml	0.64	0.43-0.86	2.67	1.79-3.54	5.94	3.98-7.89
Troponin in ng/ml	<0.5		1.77	1.19-2.36	4.98	2.85-7.10
Myoglobin in ng/ml	35.72	23.93-47.51	132.21	88.58-175.84	296.33	198.54-394.12

DOCUMENT HISTORY

PREPARED BY: *MSW* DEPT: QC VERIFIED BY: *A. Shatola* DEPT: QA
 APPROVED BY: *Fred Gibson* DEPT: Administration EFFECTIVE DATE: 2019-Oct-04
 REVISION: 0 DCO: N/A