



VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 5403(1-3) Exp date:2021-Dec-30

Acculite® CLIA	54031		54032		54033	
	Mean	Range	Mean	Range	Mean	Range
Liquichek Maternal Serum 1st Trimester						
hCG in mIU/mL	866.67	580.67-1152.67	5974.67	4003.03-7946.31	11773.83	7888.47-15659.20
hCG-XR in mIU/mL	910.0	609.70-1210.30	5782.63	3874.36-7690.89	11469.88	7684.82-15254.93
PAPP-A Mass Units in µg/ml	5.21	3.49-6.92	16.31	10.92-21.69	>32	
ffβ-hCG in ng/mL	9.04	6.06-12.02	46.76	31.33-62.19	88.72	57.64-119.79

Note 1: The HCG and HCG-XR data was obtained by diluting the sample on a 1:1000 dilution and 1:100 dilution respectively .
The dilution factor was later applied to generate the correct value for each level.

Note 2: PAPP-A data is being reported in µg/ml. To convert µg/ml into mIU/l multiply the values x 156.25

DOCUMENT HISTORY

PREPARED BY: NSW DEPT: QC

VERIFIED BY: Ashatola DEPT: QA

APPROVED BY: Zneleja DEPT: Administration

EFFECTIVE DATE: 2020-FEB-26

REVISION: 0

DCO: N/A