



VALUE ASSIGNMENT FOR BIO-RAD

MASTER LOT 6763(1-3)

Exp:2023/01/31

AccuLite® CLIA	67631		67632		67633	
	Mean	Range	Mean	Range	Mean	Range
Cardiac Markers						
CKMB in ng/ml	3.57	2.39-4.75	14.28	9.57-18.99	66.9	44.82-88.98
hsCRP in µg/ml	0.08	0.05-0.11	0.21	0.13-0.29	0.64	0.43-0.85
Myoglobin in ng/ml	71.96	48.21-95.70	126.69	84.88-168.49	302.38	202.59-402.17
Troponin I in ng/ml	<1		1.16	0.77-1.54	4.75	3.35-6.14

Note: Monobind evaluates levels 1, 2 and 3 in the Bio-Rad control.

It does not evaluate levels 1A, 1B and 1C due to minimal difference in value recognition.

DOCUMENT HISTORY

PREPARED BY: MSW DEPT: QC

VERIFIED BY: Ashatola DEPT: QA

APPROVED BY: [Signature] DEPT: Administration

EFFECTIVE DATE: 2020-03-06

REVISION: 0

DCO: N/A