



VALUE ASSIGNMENT FOR BIO-RAD

MASTER LOT 5403(1-3) Exp date:2021-Dec-30

AccuBind® ELISA	54031		54032		54033	
	Mean	Range	Mean	Range	Mean	Range
Liquichek Maternal Serum 1st Trimester						
hCG in mIU/mL	1157.90	775.79-1540.01	5436.44	3642.41-7230.46	11815.46	7916.36-15714.57
hCG-XR in mIU/mL	1077.15	721.69-1432.60	6547.98	4387.14-8708.81	11977.52	8024.94-15930.10
PAPP-A Mass Units in µg/ml	5.82	3.90-7.74	16.17	10.83-21.50	>32	
fβ-hCG in ng/ml	10.28	6.89-13.67	46.46	31.13-61.80	82.44	53.57-111.30

Note 1: The HCG and HCG-XR data was obtained by diluting the sample on a 1:1000 dilution and 1:100 dilution respectively . The dilution factor was later applied to generate the correct value for each level.

Note 2: PAPP-A data is being reported in µg/ml. To convert µg/ml into mIU/l multiply the values x 156.25

DOCUMENT HISTORY

PREPARED BY: *NYW* DEPT: QC
 APPROVED BY: *Frederick* DEPT: Administration
 REVISION: 0

VERIFIED BY: *A Shatola* DEPT: QA
 EFFECTIVE DATE: 2020-Feb-26
 DCO: N/A