



VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 5469(1-3) Exp date:2023-June-30

AccuBind® ELISA	54691		54692		54693	
	Mean	Range	Mean	Range	Mean	Range
Tumor Markers						
Ferritin in ng/ml	59.97	40.18-79.75	287.38	192.55-382.22	668.07	447.61->800
CEA Next Generation in ng/ml	2.66	1.78-3.53	25.07	16.80-33.34	82.23	55.10-109.37
CEA in ng/ml	2.94	1.97-3.91	26.82	17.97-35.67	91.94	61.60-122.29
AFP in ng/ml	8.79	5.89-11.70	84.69	56.74-112.64	263.64	176.64-350.65
PSA in ng/ml	<0.5		6.06	4.06-8.06	27.33	18.31-36.36
fPSA in ng/ml	<0.5		2.72	1.83-3.62	>10	
CA-125 in U/ml	24.87	16.66-33.08	71.43	47.86-95.0	220.81	147.94-293.68
CA-19-9 in U/ml	16.73	11.21-22.25	33.25	22.28-44.22	81.68	54.72-108.63
CA 15-3 in U/ml	27.18	18.21-36.15	53.64	35.94-71.34	92.58	62.03-123.13
PRL in ng/ml	13.42	8.99-17.85	42.86	28.71-57.00	>100	
PRL in ng/ml-HUMAN	12.45	8.34-16.56	39.63	26.55-52.70	>100	
HCG-XR in mIU/ml	3.58	2.40-4.76	66.61	44.63-88.59	309.48	207.35-411.60
HCG in mIU/ml	4.80	3.22-6.39	63.42	42.49-84.35	>200	
TG in ng/ml	3.89	2.61-5.17	22.94	15.37-30.51	52.56	32.21-69.90

PREPARED BY: <u>RSW</u> DEPT: QC	DOCUMENT HISTORY	VERIFIED BY: <u>AShatok</u> DEPT: QA
APPROVED BY: <u>Fuelgan</u> DEPT: Administration		EFFECTIVE DATE: 2020-April-20
REVISION: 0		DCO: N/A