

**VALUE ASSIGNMENT FOR BIO-RAD**  
**MASTER LOT 5469(1-3) Exp date:2023-June-30**

AccuLite® CLIA	54691		54692		54693	
	Mean	Range	Mean	Range	Mean	Range
<b>Tumor Markers</b>						
Ferritin in ng/ml	53.56	35.88-71.23	319.60	214.13-425.06	744.91	499.09->800
CEA Next Generation in ng/ml	2.06	1.38-2.74	21.76	14.58-28.94	80.91	54.21-107.61
CEA in ng/ml	2.71	1.82-3.61	30.54	20.46-40.61	100.78	67.52-134.03
AFP in ng/ml	8.15	5.46-10.84	85.88	57.54-114.22	252.68	169.30-336.07
PSA in ng/ml	<0.5		5.99	4.01-7.96	28.04	18.78-37.29
fPSA in ng/ml	<0.5		2.64	1.77-3.50	>10	
CA-125 in U/ml	21.63	14.49-28.76	76.41	51.20-101.63	202.07	135.38-268.75
CA-19-9 in U/ml	11.45	7.50-15.41	24.54	16.44-32.63	64.43	43.17-85.70
CA 15-3 in U/ml	27.83	18.65-37.02	51.28	34.36-68.20	96.16	64.43-127.89
PRL in ng/ml	12.79	8.57-17.01	43.01	28.82-57.20	>100	
HCG-XR in mIU/ml	2.77	1.86-3.68	57.95	38.82-77.07	311.29	208.56-414.02
HCG in mIU/ml	4.16	2.79-5.54	62.59	41.93-83.24	>200	
TG in ng/ml	4.52	3.03-6.01	22.20	14.87-29.52	55.31	37.06-73.56

DOCUMENT HISTORY

 PREPARED BY: RSW DEPT: QC

 VERIFIED BY: Ashatah DEPT: QA

 APPROVED BY: [Signature] DEPT: Administration

EFFECTIVE DATE: 2020-April-20

REVISION: 0

DCO: N/A