

VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT 5404(1-3) Exp date:2022-April-30						
AccuBind® ELISA	54041		54042		54043	
	Mean	Range	Mean	Range	Mean	Range
Liquichek Maternal Serum 1st Trimester						
hCG in mIU/mL	325.79	218.28-433.30	2143.34	1436.04-2850.64	5139.58	3443.52-6835.64
hCG-XR in mIU/mL	288.24	193.12-383.36	2594.96	1738.62-3451.30	5274.53	3533.93-7015.12
fβ-hCG in ng/ml	9.02	6.04-12.0	48.96	32.81-65.12	94.54	63.34-125.73

**Note 1:** The HCG and HCG-XR data was obtained by diluting Level I at 1:21 and level II and III at 1:1000 dilution respectively . The dilution factor was later applied to generate the correct value for each level.

**Note 2:** PAPP-A data is being reported in µg/ml. To convert µg/ml into mIU/l multiply the values x 156.25

## DOCUMENT HISTORY

PREPARED BY: NSW DEPT: QC  
 VERIFIED BY: A Shatola DEPT: QA  
 APPROVED BY: Fred Lopez DEPT: Administration  
 EFFECTIVE DATE: 2020-May-6  
 REVISION: 0 DCO: N/A