

VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT 5404(1-3) Exp date:2022-April-30						
Acculite® CLIA	54041		54042		54043	
	Mean	Range	Mean	Range	Mean	Range
Liquichek Maternal Serum 1st Trimester						
hCG in mIU/mL	311.15	208.47-413.83	1946.67	1238.34-2654.99	5110.0	3423.70-6796.30
hCG-XR in mIU/mL	313.64	210.14-417.13	2275.40	1524.52-3026.28	5574.83	3735.14-7414.53
fβ-hCG in ng/mL	8.33	5.58-11.08	49.40	33.10-65.71	99.83	66.89-132.77

Note 1: The HCG and HCG-XR data was obtained by diluting Level I at 1:21 and level II and III at 1:1000 dilution respectively .
The dilution factor was later applied to generate the correct value for each level.

Note 2: PAPP-A data is being reported in µg/ml. To convert µg/ml into mIU/l multiply the values x 156.25

DOCUMENT HISTORY

PREPARED BY: MSW DEPT: QC
 APPROVED BY: [Signature] DEPT: Administration
 REVISION: 0

VERIFIED BY: AShatola DEPT: QA
 EFFECTIVE DATE: 2020-May-6
 DCO: N/A