



Monobind Inc.

The World Resource for Diagnostic Products

VALUE ASSIGNMENT FOR BIO-RAD  
MASTER LOT 6029(L-3) EXP:2022/08/31

AccuLite® CLIA	6029L		60291		60292		60293	
	Mean	Range	Mean	Range	Mean	Range	Mean	Range
Liquicheck Specialty IA								
anti-TG in IU/ml	53.10	35.58-70.62	106.33	71.24-141.42	192.60	117.40-267.80	423.10	283.48-562.72
anti-TPO in IU/ml	12.25	8.21-16.29	26.58	17.81-35.34	49.98	33.48-66.47	133.53	89.46-177.59
C-peptide in ng/ml	0.83	0.55-1.10	0.88	0.59-1.17	3.56	2.39-4.74	8.64	6.27-11.02

DOCUMENT HISTORY

PREPARED BY: *MSW* DEPT: QC  
 APPROVED BY: *Falga* DEPT: Administration  
 REVISION: 0

VERIFIED BY: *Ashata* DEPT: QA  
 EFFECTIVE DATE: 2019-09-20  
 DCO: N/A