

**VALUE ASSIGNMENT FOR BIO-RAD  
MASTER LOT 4228(1-3) EXP: 2023/07/31**

AccuLite® CLIA	42281		42282		42283	
	Mean	Range	Mean	Range	Mean	Range
<b>Fertility</b>						
Ferritin in ng/ml	16.67	11.17-22.17	51.37	34.42-68.32	129.55	86.80-172.30
AFP in ng/ml	23.30	15.61-30.99	97.76	65.50-130.02	348.13	233.24-463.01
LH in mIU/ml	166.00	111.22-220.78	65.40	43.82-86.98	38.68	25.92-51.45
hCG in mIU/ml	<2		30.18	20.22-40.13	21858.75	13291.66-30425.84
Testosterone in ng/ml	0.31	0.21-0.41	3.67	2.39-4.96	8.10	4.85-11.35
Progesterone in ng/ml	18.67	12.51-24.83	43.71	29.29-58.14	>60	
E2 in pg/ml	211.80	141.91-281.69	218.52	146.41-290.63	1098.50	734.05-1462.95

DOCUMENT HISTORY

PREPARED BY:                     *NSW*                     DEPT: QC  
 APPROVED BY:                     *Fred Lopez*                     DEPT: Administration  
 REVISION: 0

VERIFIED BY:                     *Ashatoka*                     DEPT: QA  
 EFFECTIVE DATE: 2020-06-18  
 DCO: N/A