



VALUE ASSIGNMENT FOR BIO-RAD  
MASTER LOT 5405(1-3) Exp date:2022-July-31

Acculite® CLIA	54051		54052		54053	
	Mean	Range	Mean	Range	Mean	Range
Liquichek Maternal Serum 1st Trimester						
hCG in mIU/mL	1187.88	795.88-1579.87	5244.63	3513.90-6975.35	10854.50	7272.52-14436.49
hCG-XR in mIU/mL	1140.83	764.36-1517.31	5120.17	3430.51-6809.82	11798.67	7905.11-15692.23
fβ-hCG in ng/mL	9.72	5.80-13.63	45.24	30.31-60.17	100.15	67.10-133.19

**Note 1:** The HCG and HCG-XR data was obtained by diluting Level I , II and III at 1:1000 dilution respectably . The dilution factor was later applied to generate the correct value for each level.

**Note 2:** PAPP-A data is being reported in µg/ml. To convert µg/ml into mIU/l multiply the values x 156.25

DOCUMENT HISTORY

PREPARED BY: NSW DEPT: QC  
 APPROVED BY: Zoe Lapan DEPT: Administration  
 REVISION: 0

VERIFIED BY: Ashatola DEPT: QA  
 EFFECTIVE DATE: 2020-August-15  
 DCO: N/A