



VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 5405(1-3) Exp date:2022-July-31

AccuBind® ELISA	54051		54052		54053	
	Mean	Range	Mean	Range	Mean	Range
Liquichek Maternal Serum 1st Trimester						
hCG in mIU/mL	1073.50	719.25-1427.76	5017.0	3361.39-6672.61	10691.33	7163.19-14219.47
hCG-XR in mIU/mL	1010.0	599.59-1420.41	5242.20	3512.27-6972.13	10280.60	6888.0-13673.20
βhCG in ng/ml	10.05	6.73-13.36	48.42	32.44-64.40	101.98	68.33-135.63

Note 1: The HCG and HCG-XR data was obtained by diluting Level I , II and III at 1:1000 dilution respectably . The dilution factor was later applied to generate the correct value for each level.

Note 2: PAPP-A data is being reported in µg/ml. To convert µg/ml into mIU/l multiply the values x 156.25

DOCUMENT HISTORY

PREPARED BY: NSW DEPT: QC VERIFIED BY: A. Shatola DEPT: QA
APPROVED BY: Fred Johnson DEPT: Administration EFFECTIVE DATE: 2020-August-15
REVISION: 0 DCO: N/A