

VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 6764(1-3) Exp:2023/05/31

AccuLite® CLIA	67641		67642		67643	
	Mean	Range	Mean	Range	Mean	Range
Cardiac Markers						
CKMB in ng/ml	3.58	2.40-4.75	14.82	9.93-19.71	76.41	51.19-101.62
hsCRP in µg/ml	0.08	0.06-0.11	0.22	0.14-0.29	0.69	0.46-0.91
Myoglobin in ng/ml	82.7	55.41-110	139.04	93.16-184.92	260.46	174.51-346.41
Troponin I in ng/ml	<1		1.3	0.87-1.73	4.95	3.32-6.58

Note: Monobind evaluates levels 1, 2 and 3 in the Bio-Rad control.
It does not evaluate levels 1A, 1B and 1C due to minimal difference in value recognition.

DOCUMENT HISTORY

PREPARED BY: *MSW* DEPT: QC VERIFIED BY: *Ashatola* DEPT: QA
 APPROVED BY: *[Signature]* DEPT: Administration EFFECTIVE DATE: 2020-08-14
 REVISION: 0 DCO: N/A