

VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 8782(1-3) Exp:07/31/2023

AccuLite® CLIA	87821		87822		87823	
	Mean	Range	Mean	Range	Mean	Range
Cardiac Markers						
CKMB in ng/ml	1.47	0.85-2.09	12.44	8.33-16.54	64.11	42.95-85.26
hsCRP in µg/ml	0.51	0.34-0.67	2.14	1.44-2.85	5.06	3.39-6.73
Troponin in ng/ml	<0.5		2.07	1.38-2.75	6.20	4.15-8.24
Myoglobin in ng/ml	40.91	27.41-54.41	133.52	89.46-177.58	271.39	181.83-360.95

DOCUMENT HISTORY

 PREPARED BY: *Chitra Kishor* DEPT: Research

 VERIFIED BY: *MSW* DEPT: QC

 APPROVED BY: *Frederic J. ...* DEPT: Administration

EFFECTIVE DATE: 2020/10/01

REVISION: 0

DCO: N/A