

VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 9490(1-3)
Exp:2022/12/31

AccuLite® CLIA	94901		94902		94903	
	Mean	Range	Mean	Range	Mean	Range
Liquicheck Tumor Markers						
Ferritin in ng/ml	28.34	18.99-37.69	328.73	220.25-437.20	737.10	493.86->800
CEA in ng/ml	3.33	2.23-4.42	30.6	20.50-40.70	108.33	72.58-144.07
AFP in ng/ml	6.43	4.31-8.55	99.71	66.81-132.62	300.86	201.57-400.14
PSA in ng/ml	<2		9.04	6.06-12.02	40.56	27.18-53.95
fPSA in ng/ml	<0.3		4.17	2.80-5.55	>10	
CA-125 in U/ml	25.03	16.77-33.28	73.73	49.40-98.06	218.16	146.17-290.15
CA-19-9 in U/ml	4.58	3.07-6.09	18.82	12.61-25.03	61.32	41.08-81.56
CEA Next Generation in ng/ml	3.28	2.20-4.36	26.64	17.85-35.43	88	58.96-117.04
PRL in ng/ml	11.83	7.93-15.74	40.6	27.20-54.0	>100	
HCG in mIU/ml	4.01	2.69-5.34	52.14	34.94-69.35	225.74	151.24->250
TG in ng/ml	3.87	2.59-5.15	46.65	31.26-62.05	142.36	95.38-189.34
CA 15-3 in U/ml	18.91	12.67-25.15	49.96	33.48-66.45	108.31	72.57-144.06
HCG-XR in mIU/ml	<5		44.6	29.88-59.32	245.53	164.50-326.55

DOCUMENT HISTORY

 PREPARED BY: RSW DEPT: QC

 VERIFIED BY: Ashatola DEPT: QA

 APPROVED BY: Fred Jones DEPT: Administration

EFFECTIVE DATE: 2020-10-30

REVISION: 0

DCO: N/A