

VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT 9490(1-3)			Exp:2022/12/31			
AccuBind® ELISA	94901		94902		94903	
	Mean	Range	Mean	Range	Mean	Range
Liquicheck Tumor Markers						
Ferritin in ng/ml	27.62	18.51-36.73	323.24	216.57-429.91	691.71	463.45->800
CEA in ng/ml	4.51	3.02-6.0	31.76	21.28-42.25	123.84	82.97-164.71
AFP in ng/ml	7.29	4.88-9.69	92.68	62.10-123.27	311.53	208.73-414.34
PSA in ng/ml	<0.5		9.22	6.18-12.26	39.47	26.44-52.49
fPSA in ng/ml	<0.3		4.07	2.72-5.41	>10	
CA-125 in U/ml	27.54	18.45-36.62	78.19	52.38-103.99	220.94	148.03-293.85
CA-19-9 in U/ml	4.73	3.17-6.29	18.16	12.16-24.15	65.53	43.91-87.16
CEA Next Generation in ng/ml	3.46	2.32-4.60	26.91	18.03-35.78	86.09	57.68-114.49
PRL in ng/ml	12.48	8.36-16.60	39.52	26.48-52.57	>100	
HCG in mIU/ml	4.62	3.09-6.14	49.88	33.42-66.33	227.24	152.25->250
TG in ng/ml	4.2	2.81-5.59	48.7	32.63-64.77	139.57	93.51-185.63
CA 15-3 in U/ml	16.63	11.14-22.12	51.15	34.27-68.04	97.33	65.21-129.44
HCG-XR in mIU/ml	<5		50.59	33.90-67.29	236.48	158.44-314.52

## DOCUMENT HISTORY

 PREPARED BY: RSU DEPT: QC

 VERIFIED BY: Ashoka DEPT: QA

 APPROVED BY: Frailapan DEPT: Administration

EFFECTIVE DATE: 2020-10-30

REVISION: 0

DCO: N/A