



VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 5406(1-3) Exp date: 2022-09-30

Acculite® CLIA	54061		54062		54063	
	Mean	Range	Mean	Range	Mean	Range
Liquichek Maternal Serum 1st Trimester						
hCG in mIU/mL	1050.0	703.50-1396.50	4833.33	3238.33-6428.33	9850.0	6599.50-13100.50
hCG-XR in mIU/mL	850.71	569.98-1131.45	4511.86	3022.94-6000.77	10063.57	6742.59-13384.55
fβ-hCG in ng/mL	10.74	7.19-14.28	48.22	32.31-64.14	101.99	68.34-135.65
PAPP-A in µg/mL	2.98	2.0-3.96	10.72	7.18-14.25	>32	

Note 1: The HCG data was obtained by diluting Level I , II and III at 1:1000 dilution while the HCG-XR data was obtained by diluting Level I , II and III at 1:1000 dilution respectively . The dilution factor was later applied to generate the correct value for each level.

Note 2: PAPP-A data is being reported in µg/ml. To convert µg/ml into mIU/l multiply the values x 156.25

DOCUMENT HISTORY

PREPARED BY: NSW DEPT: QC VERIFIED BY: Ashatola DEPT: QA
 APPROVED BY: [Signature] DEPT: Administration EFFECTIVE DATE: 2020-Oct-30
 REVISION: 0 DCO: N/A