



VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 5406(1-3) Exp date: 2022-09-30

AccuBind® ELISA	54061		54062		54063	
	Mean	Range	Mean	Range	Mean	Range
Liquichek Maternal Serum 1st Trimester						
hCG in mIU/mL	981.33	657.49-1305.17	4804.33	3218.90-6389.76	9910.17	6639.81-13180.52
hCG-XR in mIU/mL	830.64	556.53-1104.75	4946.54	3314.18-6578.89	10934.06	7325.82-14542.30
fβ-hCG in ng/ml	11.95	8.01-15.89	50.82	34.05-67.59	102.25	68.51-135.99
PAPP-A in µg/ml	3.08	2.07-4.10	9.03	6.05-12.01	>32	

Note 1: The HCG data was obtained by diluting Level I , II and III at 1:1000 dilution while the HCG-XR data was obtained by diluting Level I , II and III at 1:1000 dilution respectably . The dilution factor was later applied to generate the correct value for each level.

Note 2: PAPP-A data is being reported in µg/ml. To convert µg/ml into mIU/l multiply the values x 156.25

DOCUMENT HISTORY

PREPARED BY: *NSP* DEPT: QC VERIFIED BY: *AShatola* DEPT: QA
 APPROVED BY: *Fred Lopez* DEPT: Administration EFFECTIVE DATE: 2020-Oct-30
 REVISION: 0 DCO: N/A