



VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 6765(1-3) Exp:2023/11/30

AccuLite® CLIA	67651		67652		67653	
	Mean	Range	Mean	Range	Mean	Range
Cardiac Markers						
CKMB in ng/ml	3.54	1.66-5.41	15.26	10.22-20.29	63.95	42.85-85.05
hsCRP in µg/ml	0.1	0.07-0.13	0.23	0.15-0.30	0.59	0.40-0.79
Myoglobin in ng/ml	61.77	41.39-82.16	101.77	68.19-135.35	209.49	140.36-278.62
Troponin I in ng/ml	<1		1.32	0.89-1.76	4.22	2.82-5.61

Note: Monobind evaluates levels 1, 2 and 3 in the Bio-Rad control.
It does not evaluate levels 1A, 1B and 1C due to minimal difference in value recognition.

DOCUMENT HISTORY

PREPARED BY: MSW DEPT: QC VERIFIED BY: Ashatola DEPT: QA
APPROVED BY: [Signature] DEPT: Administration EFFECTIVE DATE: 2021-01-08
REVISION: 0 DCO: N/A