

VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT 9491(1-3)			Exp:2023/04/30			
AccuBind® ELISA	94911		94912		94913	
	Mean	Range	Mean	Range	Mean	Range
Liquicheck Tumor Markers						
Ferritin in ng/ml	27.89	18.69-37.10	297.38	199.24-395.51	665.25	445.72->800
CEA in ng/ml	3.13	1.97-4.29	30.78	20.62-40.93	124.41	83.35-165.47
AFP in ng/ml	7.81	5.23-10.38	104.49	70.01-138.97	383.3	256.81-509.78
PSA in ng/ml	<0.5		9.16	6.14-12.18	40.57	27.18-53.96
fPSA in ng/ml	<0.3		3.97	2.66-5.28	>10	
CA-125 in U/ml	26.98	18.07-35.88	73.78	49.43-98.13	202.2	135.47-268.92
CA-19-9 in U/ml	4.75	3.19-6.32	17.93	12.01-23.84	69.59	46.63-92.56
CEA Next Generation in ng/ml	<5		27.07	18.14-36.01	88.66	59.40-117.92
PRL in ng/ml	12.42	8.32-16.52	41.5	27.80-55.19	>100	
HCG in mIU/ml	4.72	3.16-6.27	57.15	38.29-76.01	223.51	149.75->250
TG in ng/ml	4.32	2.90-5.75	52.73	35.33-70.14	140.13	93.88-186.37
CA 15-3 in U/ml	19.59	13.12-26.05	51.85	34.74-68.97	98.24	65.82-130.66
HCG-XR in mIU/ml	<5		53.02	35.52-70.52	242.64	162.57-322.71

## DOCUMENT HISTORY

 PREPARED BY: RSW DEPT: QC

 VERIFIED BY: Ashoka DEPT: QA

 APPROVED BY: Frailapan DEPT: Administration

EFFECTIVE DATE: 2021-03-12

REVISION: 0

DCO: N/A