



VALUE ASSIGNMENT FOR BIO-RAD

MASTER LOT 5407(1-3) Exp date: 2023-05-31

Acculite® CLIA	54071		54072		54073	
	Mean	Range	Mean	Range	Mean	Range
Liquichek Maternal Serum 1st Trimester						
hCG in mIU/mL	981.22	657.42-1305.02	5241.10	3511.54-6970.66	9998.80	6699.19-13298.40
hCG-XR in mIU/mL	941.66	630.91-1252.41	5051.23	3384.32-6718.13	10139.90	6793.74-13486.07
fβ-hCG in ng/mL	9.95	6.66-13.23	46.06	30.86-61.26	91.94	61.60-122.28
PAPP-A in µg/mL	3.16	2.11-4.20	9.80	6.56-13.03	>32	

Note 1: The HCG data was obtained by diluting Level I , II and III at 1:1000 dilution while the HCG-XR data was obtained by diluting Level I , II and III at 1:100 dilution respectably . The dilution factor was later applied to generate the correct value for each level.

Note 2: PAPP-A data is being reported in µg/ml. To convert µg/ml into mIU/l multiply the values x 156.25

DOCUMENT HISTORY

PREPARED BY: MSW DEPT: QC VERIFIED BY: Ashatola DEPT: QA
 APPROVED BY: [Signature] DEPT: Administration EFFECTIVE DATE: 2021-JUL-01
 REVISION: 0 DCO: N/A