



VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 5407(1-3) Exp date: 2023-05-31

AccuBind® ELISA	54071		54072		54073	
	Mean	Range	Mean	Range	Mean	Range
Liquichek Maternal Serum 1st Trimester						
hCG in mIU/mL	949.56	636.20-1262.91	5163.70	3459.68-6867.72	9902.36	6634.58-13170.14
hCG-XR in mIU/mL	949.70	636.30-1263.10	5140.06	3443.84-6836.28	9458.95	6337.50-12580.41
fβ-hCG in ng/ml	10.51	7.04-13.98	47.67	31.94-63.39	92.98	62.30-123.67
PAPP-A in µg/ml	3.21	2.15-4.28	9.98	6.69-13.27	>32	

Note 1: The HCG data was obtained by diluting Level I , II and III at 1:1000 dilution while the HCG-XR data was obtained by diluting Level I , II and III at 1:100 dilution respectively . The dilution factor was later applied to generate the correct value for each level.

Note 2: PAPP-A data is being reported in µg/ml. To convert µg/ml into mIU/l multiply the values x 156.25

DOCUMENT HISTORY

PREPARED BY: *NSP* DEPT: QC VERIFIED BY: *AShatola* DEPT: QA
APPROVED BY: *Fred Lopez* DEPT: Administration EFFECTIVE DATE: 2021-JUL-01
REVISION: 0 DCO: N/A