

VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT 9492(1-3)			Exp:2023/12/31			
AccuLite® CLIA	94921		94922		94923	
	Mean	Range	Mean	Range	Mean	Range
Liquicheck Tumor Markers						
Ferritin in ng/ml	28.08	18.81-37.35	376.51	320.68-432.34	>800	
CEA in ng/ml	3.7	2.48-4.92	27.75	18.59-36.91	102.5	68.68-136.33
AFP in ng/ml	7.88	5.28-10.47	108.2	72.49-143.91	313.8	210.25-417.35
PSA in ng/ml	0.28	0.19-0.37	10.2	6.83-13.57	45.05	30.18-59.92
fPSA in ng/ml	<0.5		3.82	2.56-5.08	>10	
CA-125 in U/ml	21.17	14.18-28.15	62.06	41.58-82.53	196.14	131.41-260.86
CA-19-9 in U/ml	3.7	2.48-4.92	20.42	13.68-27.16	79.27	53.11-105.43
CEA Next Generation in ng/ml	<5		25.6	17.15-34.05	89.83	60.18-119.47
PRL in ng/ml	13.98	9.36-18.59	50.68	33.95-67.40	>100	
HCG in mIU/ml	4.8	3.22-6.38	43.95	29.45-58.45	241.6	161.87->250
TG in ng/ml	4.98	3.34-6.62	47.75	31.99-63.50	150.07	100.55-199.59
CA 15-3 in U/ml	19.03	12.68-25.39	77.51	51.93-103.08	170.36	106.74-233.97
HCG-XR in mIU/ml	5.54	3.71-7.36	59.3	39.73-78.87	232.23	155.60-308.87

DOCUMENT HISTORY			
PREPARED BY: <u>MSW</u>	DEPT: QC	VERIFIED BY: <u>Ashatoka</u>	DEPT: QA
APPROVED BY: <u>Fred Lopez</u>	DEPT: Administration	EFFECTIVE DATE:	2021/10/26
REVISION: 0		DCO: N/A	